


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN120720\
Data File : VN064934.D
Acq On : 7 Dec 2020 9:57
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Title : SW846 8260
Last Update : Mon Nov 23 13:54:12 2020



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3294

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	6841	PASS
75	95	30	60	46.4	20098	PASS
95	95	100	100	100.0	43272	PASS
96	95	5	9	7.1	3078	PASS
173	174	0.00	2	1.4	473	PASS
174	95	50	100	80.8	34960	PASS
175	174	5	9	7.4	2588	PASS
176	174	95	101	96.8	33848	PASS
177	176	5	9	6.1	2074	PASS