

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069810.D
 Acq On : 7 Dec 2021 13:17
 Operator : JC/MD
 Sample : M4956-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-20211203

Quant Time: Dec 08 04:34:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1377076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2459005	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2585583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	974246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1056360	49.683	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.360%		
35) Dibromofluoromethane	8.021	113	722458	46.905	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.800%		
50) Toluene-d8	10.443	98	3138123	50.740	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.480%		
62) 4-Bromofluorobenzene	12.731	95	1194168	53.297	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	106.600%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	83046	6.493	ug/l	92
20) Methylene Chloride	5.095	84	14670	0.863	ug/l	91
25) 2-Butanone	7.353	43	16463	0.947	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069810.D
Acq On : 7 Dec 2021 13:17
Operator : JC/MD
Sample : M4956-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-20211203

Quant Time: Dec 08 04:34:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

