

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069815.D
 Acq On : 7 Dec 2021 15:24
 Operator : JC/MD
 Sample : M4956-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE123D1-20211204

Quant Time: Dec 08 04:36:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/08/2021
 Supervised By : Amit Patel 12/14/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1335219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2393101	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2579615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1004862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1016783	49.321	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.640%		
35) Dibromofluoromethane	8.024	113	711275	47.450	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.900%		
50) Toluene-d8	10.440	98	3084351	51.244	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.480%		
62) 4-Bromofluorobenzene	12.731	95	1219682	55.935	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	111.880%		
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.200	101	9533m	0.722	ug/l	
16) Acetone	4.307	43	19377m	1.562	ug/l	
44) Trichloroethene	9.217	130	26322	1.584	ug/l	90
64) Tetrachloroethene	10.974	164	5311	0.304	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069815.D
Acq On : 7 Dec 2021 15:24
Operator : JC/MD
Sample : M4956-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE123D1-20211204

Quant Time: Dec 08 04:36:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/08/2021
Supervised By : Amit Patel 12/14/2021

