

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069817.D
 Acq On : 7 Dec 2021 16:15
 Operator : JC/MD
 Sample : M4956-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE123D2-20211204

Quant Time: Dec 08 04:36:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1293562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2297457	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2430642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	923851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	984302	49.283	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.560%
35) Dibromofluoromethane	8.021	113	683501	47.496	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.000%
50) Toluene-d8	10.440	98	2957252	51.178	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.360%
62) 4-Bromofluorobenzene	12.731	95	1138463	54.384	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.760%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	30127	2.508	ug/l	92
44) Trichloroethene	9.217	130	27033	1.694	ug/l	84
64) Tetrachloroethene	10.979	164	19142	1.161	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069817.D
Acq On : 7 Dec 2021 16:15
Operator : JC/MD
Sample : M4956-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE123D2-20211204

Quant Time: Dec 08 04:36:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

