

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069822.D
 Acq On : 7 Dec 2021 18:23
 Operator : JC/MD
 Sample : M4956-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW201D-20211205

Quant Time: Dec 08 04:38:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 12/20/2021
 Supervised By : Mahesh Dadoda 12/21/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1331216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2387804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2530925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	980516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1015130	49.388	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.780%		
35) Dibromofluoromethane	8.021	113	714316	47.759	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.520%		
50) Toluene-d8	10.443	98	3074910	51.201	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.400%		
62) 4-Bromofluorobenzene	12.731	95	1198323	55.078	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	110.160%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.073	85	108518	6.989	ug/l	98
7) Trichlorofluoromethane	3.381	101	2538321	108.130	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.210	101	25703	1.953	ug/l	93
12) 1,1-Dichloroethene	4.183	96	404575	31.906	ug/l	87
24) 1,1-Dichloroethane	6.391	63	78643	2.718	ug/l	98
27) cis-1,2-Dichloroethene	7.332	96	29734	1.787	ug/l	95
30) Chloroform	7.820	83	94822	3.245	ug/l	93
32) 1,1,1-Trichloroethane	8.013	97	93541m	3.759	ug/l	
44) Trichloroethene	9.217	130	992824	59.868	ug/l	92
64) Tetrachloroethene	10.977	164	832149	48.490	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069822.D
Acq On : 7 Dec 2021 18:23
Operator : JC/MD
Sample : M4956-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW201D-20211205

Quant Time: Dec 08 04:38:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/20/2021
Supervised By : Mahesh Dadoda 12/21/2021

