

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069823.D
 Acq On : 7 Dec 2021 18:48
 Operator : JC/MD
 Sample : M4956-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW201D1-20211205

Quant Time: Dec 08 04:38:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1289588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2323487	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2419235	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	892161	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	992746	49.86	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.72%
35) Dibromofluoromethane	8.024	113	692432	47.58	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.16%
50) Toluene-d8	10.443	98	2950139	50.48	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.96%
62) 4-Bromofluorobenzene	12.731	95	1125102	53.14	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.28%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.215	130	16861	1.04	ug/l	99
64) Tetrachloroethene	10.979	164	6125	0.37	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069823.D
Acq On : 7 Dec 2021 18:48
Operator : JC/MD
Sample : M4956-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW201D1-20211205

Quant Time: Dec 08 04:38:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

