

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069839.D
 Acq On : 8 Dec 2021 2:51
 Operator : JC/MD
 Sample : M4957-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW179D1-20211206

Quant Time: Dec 08 06:37:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1304974	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2337117	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2493475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	945289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1007969	50.026	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.060%	
35) Dibromofluoromethane	8.021	113	695876	47.535	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 95.080%	
50) Toluene-d8	10.440	98	3037197	51.670	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.340%	
62) 4-Bromofluorobenzene	12.731	95	1155393	54.256	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.520%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.327	96	148091	9.079	ug/l	96
44) Trichloroethene	9.217	130	1227907	75.650	ug/l	91
64) Tetrachloroethene	10.977	164	109383	6.470	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069839.D
Acq On : 8 Dec 2021 2:51
Operator : JC/MD
Sample : M4957-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW179D1-20211206

Quant Time: Dec 08 06:37:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

