

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069842.D
 Acq On : 8 Dec 2021 4:07
 Operator : JC/MD
 Sample : M4957-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE125D3-20211206

Quant Time: Dec 08 06:38:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 12/08/2021
 Supervised By : Amit Patel 12/14/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1305519	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2350472	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2513335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	923641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1021720	50.688	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.380%			
35) Dibromofluoromethane	8.024	113	696875	47.333	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 94.660%			
50) Toluene-d8	10.440	98	3050746	51.605	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.220%			
62) 4-Bromofluorobenzene	12.731	95	1158435	54.090	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 108.180%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.073	85	4004	0.263	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.213	101	761710	59.025	ug/l	96
12) 1,1-Dichloroethene	4.191	96	9928m	0.798	ug/l	
27) cis-1,2-Dichloroethene	7.327	96	20585	1.261	ug/l	92
30) Chloroform	7.817	83	6919	0.241	ug/l	88
38) Carbon Tetrachloride	8.209	117	6835	0.312	ug/l	95
44) Trichloroethene	9.217	130	2075541	127.145	ug/l	92
60) Dibromochloromethane	10.977	129	68416	5.986	ug/l #	11
64) Tetrachloroethene	10.977	164	71738	4.209	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069842.D
Acq On : 8 Dec 2021 4:07
Operator : JC/MD
Sample : M4957-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE125D3-20211206

Quant Time: Dec 08 06:38:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/08/2021
Supervised By :Amit Patel 12/14/2021

