

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069844.D
 Acq On : 8 Dec 2021 4:58
 Operator : JC/MD
 Sample : M4956-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW202D1-20211204

Quant Time: Dec 08 06:39:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1282389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2315321	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2517289	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	949449	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1006378	50.83	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.66%
35) Dibromofluoromethane	8.021	113	694596	47.89	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.78%
50) Toluene-d8	10.440	98	3013922	51.76	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.52%
62) 4-Bromofluorobenzene	12.731	95	1181639	56.01	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.02%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.220	130	13638	0.85	ug/l	91
64) Tetrachloroethene	10.974	164	5015	0.29	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069844.D
Acq On : 8 Dec 2021 4:58
Operator : JC/MD
Sample : M4956-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW202D1-20211204

Quant Time: Dec 08 06:39:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

