

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069827.D
 Acq On : 7 Dec 2021 20:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 08 04:40:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.583	0.582	0.2	88	0.00
3 P	Chloromethane	0.731	0.677	7.4	88	0.00
4 C	Vinyl Chloride	0.672	0.649	3.4#	89	0.00
5 T	Bromomethane	0.365	0.371	-1.6	93	0.00
6 T	Chloroethane	0.412	0.398	3.4	90	0.00
7 T	Trichlorofluoromethane	0.882	0.840	4.8	89	0.00
8 T	Diethyl Ether	0.340	0.334	1.8	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.492	0.4	93	0.00
10 T	Methyl Iodide	0.675	0.670	0.7	87	0.00
11 T	Tert butyl alcohol	0.146	0.119	18.5	75	0.00
12 CM	1,1-Dichloroethene	0.476	0.466	2.1#	91	0.00
13 T	Acrolein	0.124	0.159	-28.2#	126	0.00
14 T	Allyl chloride	1.073	1.064	0.8	92	0.00
15 T	Acrylonitrile	0.393	0.363	7.6	84	0.00
16 T	Acetone	0.464	0.348	25.0#	68	0.00
17 T	Carbon Disulfide	1.230	1.207	1.9	89	0.00
18 T	Methyl Acetate	1.402	1.288	8.1	87	0.00
19 T	Methyl tert-butyl Ether	1.852	1.850	0.1	92	0.00
20 T	Methylene Chloride	0.617	0.584	5.3	93	0.00
21 T	trans-1,2-Dichloroethene	0.517	0.505	2.3	91	0.00
22 T	Diisopropyl ether	2.098	2.157	-2.8	93	0.00
23 T	Vinyl Acetate	1.707	1.770	-3.7	91	0.00
24 P	1,1-Dichloroethane	1.087	1.096	-0.8	93	0.00
25 T	2-Butanone	0.631	0.538	14.7	77	0.00
26 T	2,2-Dichloropropane	0.845	0.823	2.6	91	0.00
27 T	cis-1,2-Dichloroethene	0.625	0.612	2.1	90	0.00
28 T	Bromochloromethane	0.565	0.547	3.2	90	0.00
29 T	Tetrahydrofuran	0.383	0.348	9.1	81	0.00
30 C	Chloroform	1.097	1.091	0.5#	92	0.00
31 T	Cyclohexane	1.074	1.029	4.2	91	0.00
32 T	1,1,1-Trichloroethane	0.935	0.939	-0.4	91	0.00
33 S	1,2-Dichloroethane-d4	0.772	0.726	6.0	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.313	0.301	3.8	88	0.00
36 T	1,1-Dichloropropene	0.476	0.480	-0.8	91	0.00
37 T	Ethyl Acetate	0.708	0.658	7.1	83	0.00
38 T	Carbon Tetrachloride	0.467	0.476	-1.9	91	0.00
39 T	Methylcyclohexane	0.566	0.571	-0.9	89	0.00
40 TM	Benzene	1.455	1.459	-0.3	91	0.00
41 T	Methacrylonitrile	0.333	0.327	1.8	86	0.00
42 TM	1,2-Dichloroethane	0.587	0.578	1.5	90	0.00
43 T	Isopropyl Acetate	1.048	1.031	1.6	88	0.00
44 TM	Trichloroethene	0.347	0.353	-1.7	91	0.00
45 C	1,2-Dichloropropane	0.391	0.397	-1.5#	93	0.00
46 T	Dibromomethane	0.256	0.257	-0.4	91	0.00
47 T	Bromodichloromethane	0.485	0.512	-5.6	93	0.00
48 T	Methyl methacrylate	0.489	0.522	-6.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069827.D
 Acq On : 7 Dec 2021 20:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 08 04:40:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.007	22.2	71	0.00
50 S	Toluene-d8	1.258	1.199	4.7	87	0.00
51 T	4-Methyl-2-Pentanone	0.680	0.671	1.3	84	0.00
52 CM	Toluene	0.888	0.909	-2.4#	91	0.00
53 T	t-1,3-Dichloropropene	0.510	0.548	-7.5	91	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.596	-8.8	91	0.00
55 T	1,1,2-Trichloroethane	0.358	0.362	-1.1	90	0.00
56 T	Ethyl methacrylate	0.608	0.626	-3.0	89	0.00
57 T	1,3-Dichloropropane	0.631	0.640	-1.4	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.154	0.139	9.7	77	0.00
59 T	2-Hexanone	0.514	0.477	7.2	79	0.00
60 T	Dibromochloromethane	0.340	0.364	-7.1	91	0.00
61 T	1,2-Dibromoethane	0.371	0.375	-1.1	89	0.00
62 S	4-Bromofluorobenzene	0.456	0.434	4.8	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.339	0.334	1.5	89	0.00
65 PM	Chlorobenzene	1.014	1.027	-1.3	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.376	-5.3	91	0.00
67 C	Ethyl Benzene	1.843	1.911	-3.7#	90	0.00
68 T	m/p-Xylenes	0.671	0.697	-3.9	90	0.00
69 T	o-Xylene	0.676	0.698	-3.3	90	0.00
70 T	Styrene	1.072	1.141	-6.4	90	0.00
71 P	Bromoform	0.264	0.284	-7.6	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	4.449	4.384	1.5	90	0.00
74 T	N-amyl acetate	1.967	1.909	2.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.538	1.436	6.6	87	0.00
76 T	1,2,3-Trichloropropane	1.434	1.312	8.5	84	0.00
77 T	Bromobenzene	1.045	1.001	4.2	89	0.00
78 T	n-propylbenzene	4.950	4.983	-0.7	90	0.00
79 T	2-Chlorotoluene	3.128	3.075	1.7	91	0.00
80 T	1,3,5-Trimethylbenzene	3.581	3.572	0.3	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.396	0.383	3.3	81	0.00
82 T	4-Chlorotoluene	2.942	2.913	1.0	90	0.00
83 T	tert-Butylbenzene	3.171	3.119	1.6	89	0.00
84 T	1,2,4-Trimethylbenzene	3.410	3.517	-3.1	90	0.00
85 T	sec-Butylbenzene	4.301	4.295	0.1	88	0.00
86 T	p-Isopropyltoluene	3.362	3.446	-2.5	89	0.00
87 T	1,3-Dichlorobenzene	1.717	1.675	2.4	88	0.00
88 T	1,4-Dichlorobenzene	1.687	1.628	3.5	89	0.00
89 T	n-Butylbenzene	2.727	2.784	-2.1	88	0.00
90 T	Hexachloroethane	0.590	0.614	-4.1	90	0.00
91 T	1,2-Dichlorobenzene	1.673	1.622	3.0	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.278	0.259	6.8	82	0.00
93 T	1,2,4-Trichlorobenzene	0.760	0.767	-0.9	90	0.00
94 T	Hexachlorobutadiene	0.443	0.420	5.2	84	0.00
95 T	Naphthalene	2.199	2.174	1.1	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
Data File : VN069827.D
Acq On : 7 Dec 2021 20:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 08 04:40:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.742	0.9	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6