


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120721\
 Data File : VN069828.D
 Acq On : 7 Dec 2021 21:21
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Title : SW846 8260
 Last Update : Fri Dec 03 08:46:47 2021



AutoFind: Scans 4004, 4005, 4006; Background Corrected with Scan 3987

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.3	48149	PASS
75	95	30	60	56.3	116483	PASS
95	95	100	100	100.0	206997	PASS
96	95	5	9	6.5	13552	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.3	155968	PASS
175	174	5	9	7.4	11591	PASS
176	174	95	101	96.8	150997	PASS
177	176	5	9	6.2	9428	PASS