

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064964.D
 Acq On : 8 Dec 2020 13:01
 Operator : JC/MD
 Sample : PB133456ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#01

Quant Time: Dec 09 07:07:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	175307	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	313001	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	343776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	143900	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	107828	52.91	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	105.82%		
35) Dibromofluoromethane	7.55	113	101148	54.01	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	108.02%		
50) Toluene-d8	10.06	98	410350	53.62	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.24%		
62) 4-Bromofluorobenzene	12.38	95	158020	57.70	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	115.40%		

Target Compounds

					Qvalue
16) Acetone	3.78	43	5905	8.814 ug/l	97
20) Methylene Chloride	4.50	84	8693	4.141 ug/l	91
43) Isopropyl Acetate	8.13	43	15992	3.556 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064964.D
Acq On : 8 Dec 2020 13:01
Operator : JC/MD
Sample : PB133456ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#01

Quant Time: Dec 09 07:07:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

