

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064967.D
 Acq On : 8 Dec 2020 14:13
 Operator : JC/MD
 Sample : PB133456ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#04

Quant Time: Dec 09 07:14:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	172919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	321569	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	338395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	143931	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	108737	54.09	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	108.18%		
35) Dibromofluoromethane	7.55	113	101073	52.53	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	105.06%		
50) Toluene-d8	10.06	98	405258	51.55	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.10%		
62) 4-Bromofluorobenzene	12.38	95	154144	54.78	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	109.56%		

Target Compounds

					Qvalue
16) Acetone	3.78	43	5692	8.613 ug/l	92
20) Methylene Chloride	4.51	84	4924	2.378 ug/l	94
43) Isopropyl Acetate	8.13	43	16813	3.639 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064967.D
Acq On : 8 Dec 2020 14:13
Operator : JC/MD
Sample : PB133456ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#04

Quant Time: Dec 09 07:14:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

