

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064969.D
 Acq On : 8 Dec 2020 15:01
 Operator : JC/MD
 Sample : PB133456ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#06

Quant Time: Dec 09 07:17:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	170559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	324186	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	337716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	142909	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.98	65	107258	54.09	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.18%
35) Dibromofluoromethane	7.55	113	99923	51.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.02%
50) Toluene-d8	10.06	98	403882	50.96	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.92%
62) 4-Bromofluorobenzene	12.38	95	154218	54.37	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.74%

Target Compounds

					Qvalue	
16) Acetone	3.78	43	5343	8.197	ug/l	95
20) Methylene Chloride	4.51	84	5750	2.815	ug/l #	83
43) Isopropyl Acetate	8.13	43	17248	3.703	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064969.D
Acq On : 8 Dec 2020 15:01
Operator : JC/MD
Sample : PB133456ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#06

Quant Time: Dec 09 07:17:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

