

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064970.D
 Acq On : 8 Dec 2020 15:25
 Operator : JC/MD
 Sample : L4599-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-TB-20201203

Quant Time: Dec 08 16:25:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	166247	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	313156	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	318787	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	137488	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	105146	54.40	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	108.80%	
35) Dibromofluoromethane	7.55	113	97653	52.11	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	104.22%	
50) Toluene-d8	10.06	98	382369	49.94	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	99.88%	
62) 4-Bromofluorobenzene	12.38	95	143497	52.37	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	104.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064970.D
Acq On : 8 Dec 2020 15:25
Operator : JC/MD
Sample : L4599-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-TB-20201203

Quant Time: Dec 08 16:25:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

