

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064972.D
 Acq On : 8 Dec 2020 16:13
 Operator : JC/MD
 Sample : PB133456ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#09

Quant Time: Dec 09 07:19:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	172590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	324895	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	339956	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	145957	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.98	65	109056	54.35	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.70%	
35) Dibromofluoromethane		7.55	113	100327	51.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.22%	
50) Toluene-d8		10.06	98	409620	51.57	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.14%	
62) 4-Bromofluorobenzene		12.38	95	156713	55.13	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.26%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	5441	8.249	ug/l #	88
20) Methylene Chloride	4.50	84	4932	2.386	ug/l	88
43) Isopropyl Acetate	8.13	43	19056	4.082	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064972.D
Acq On : 8 Dec 2020 16:13
Operator : JC/MD
Sample : PB133456ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#09

Quant Time: Dec 09 07:19:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

