

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064973.D
 Acq On : 8 Dec 2020 16:37
 Operator : JC/MD
 Sample : PB133456ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#10

Quant Time: Dec 09 07:20:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	172218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	326851	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	342587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	140275	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	108742	54.31	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.62%	
35) Dibromofluoromethane		7.55	113	100340	51.30	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.60%	
50) Toluene-d8		10.06	98	416402	52.11	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.22%	
62) 4-Bromofluorobenzene		12.37	95	153861	53.80	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	6012	9.135	ug/l #	80
20) Methylene Chloride	4.50	84	5258	2.549	ug/l	95
43) Isopropyl Acetate	8.13	43	18385	3.915	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064973.D
Acq On : 8 Dec 2020 16:37
Operator : JC/MD
Sample : PB133456ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#10

Quant Time: Dec 09 07:20:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

