

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064975.D
 Acq On : 8 Dec 2020 17:26
 Operator : JC/MD
 Sample : PB133456ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#12

Quant Time: Dec 09 07:23:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	167492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	325216	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	339975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	139317	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	109401	56.18	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	112.36%		
35) Dibromofluoromethane	7.55	113	100357	51.57	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	103.14%		
50) Toluene-d8	10.06	98	407748	51.28	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.56%		
62) 4-Bromofluorobenzene	12.38	95	152413	53.56	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	107.12%		

Target Compounds

					Qvalue
16) Acetone	3.79	43	5578	8.714 ug/l	95
20) Methylene Chloride	4.51	84	5579	2.781 ug/l	90
43) Isopropyl Acetate	8.13	43	17884	3.828 ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064975.D
Acq On : 8 Dec 2020 17:26
Operator : JC/MD
Sample : PB133456ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#12

Quant Time: Dec 09 07:23:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

