

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064978.D
 Acq On : 8 Dec 2020 18:38
 Operator : JC/MD
 Sample : PB133456ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#15

Quant Time: Dec 09 07:29:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	171881	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	333284	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	346279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	142713	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	109655	54.88	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.76%	
35) Dibromofluoromethane		7.55	113	100840	50.57	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.14%	
50) Toluene-d8		10.06	98	412698	50.65	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.30%	
62) 4-Bromofluorobenzene		12.38	95	155088	53.18	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	6211	9.455	ug/l	95
20) Methylene Chloride	4.50	84	9320	4.528	ug/l	97
43) Isopropyl Acetate	8.13	43	19342	4.039	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064978.D
Acq On : 8 Dec 2020 18:38
Operator : JC/MD
Sample : PB133456ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#15

Quant Time: Dec 09 07:29:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

