

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064979.D
 Acq On : 8 Dec 2020 19:02
 Operator : JC/MD
 Sample : PB133456ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#16

Quant Time: Dec 09 07:31:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	164691	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	315747	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	328810	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	138995	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	103551	54.08	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.16%	
35) Dibromofluoromethane		7.55	113	96547	51.10	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.20%	
50) Toluene-d8		10.06	98	388813	50.37	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.74%	
62) 4-Bromofluorobenzene		12.37	95	147369	53.34	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	5468	8.688	ug/l	99
20) Methylene Chloride	4.50	84	5262	2.668	ug/l	94
43) Isopropyl Acetate	8.13	43	18096	3.989	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064979.D
Acq On : 8 Dec 2020 19:02
Operator : JC/MD
Sample : PB133456ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#16

Quant Time: Dec 09 07:31:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

