

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064981.D
 Acq On : 8 Dec 2020 19:50
 Operator : JC/MD
 Sample : PB133456ZHE#18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#18

Quant Time: Dec 09 07:33:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	169601	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	332545	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	345401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	146657	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	108919	55.24	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.48%	
35) Dibromofluoromethane		7.55	113	101565	51.04	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.08%	
50) Toluene-d8		10.06	98	415661	51.12	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.24%	
62) 4-Bromofluorobenzene		12.38	95	154763	53.19	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	5813	8.968	ug/l	92
20) Methylene Chloride	4.50	84	5748	2.830	ug/l #	87
43) Isopropyl Acetate	8.13	43	17995	3.766	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064981.D
Acq On : 8 Dec 2020 19:50
Operator : JC/MD
Sample : PB133456ZHE#18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#18

Quant Time: Dec 09 07:33:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

