

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064986.D
 Acq On : 8 Dec 2020 23:02
 Operator : JC/MD
 Sample : VN1208WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1208WBL02

Quant Time: Dec 09 08:05:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	170068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	328381	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	323696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	141944	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	108745	55.00	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.00%
35) Dibromofluoromethane	7.55	113	101459	51.64	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.28%
50) Toluene-d8	10.06	98	397818	49.55	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.10%
62) 4-Bromofluorobenzene	12.38	95	150451	52.36	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064986.D
Acq On : 8 Dec 2020 23:02
Operator : JC/MD
Sample : VN1208WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1208WBL02

Quant Time: Dec 09 08:05:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

