

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064989.D
 Acq On : 9 Dec 2020 00:14
 Operator : JC/MD
 Sample : PB133456ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#07

Quant Time: Dec 09 08:10:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	158163	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	290143	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	305467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	129057	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	99177	53.94	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	107.88%		
35) Dibromofluoromethane	7.55	113	93725	53.99	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	107.98%		
50) Toluene-d8	10.06	98	371074	52.31	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.62%		
62) 4-Bromofluorobenzene	12.38	95	137172	54.03	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	108.06%		

Target Compounds

					Qvalue
16) Acetone	3.78	43	5956	9.854 ug/l	97
20) Methylene Chloride	4.50	84	6444	3.402 ug/l	84
43) Isopropyl Acetate	8.13	43	16420	3.939 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064989.D
Acq On : 9 Dec 2020 00:14
Operator : JC/MD
Sample : PB133456ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#07

Quant Time: Dec 09 08:10:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

