

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064992.D
 Acq On : 9 Dec 2020 1:27
 Operator : JC/MD
 Sample : L4965-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT014

Quant Time: Dec 09 08:15:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	170056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	327821	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	344903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	144310	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	110580	55.93	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.86%	
35) Dibromofluoromethane		7.55	113	101592	51.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.58%	
50) Toluene-d8		10.06	98	413949	51.65	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.30%	
62) 4-Bromofluorobenzene		12.38	95	153534	53.53	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.79	43	6173	9.498	ug/l	94
20) Methylene Chloride	4.51	84	8759	4.301	ug/l	90
43) Isopropyl Acetate	8.13	43	16695	3.545	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064992.D
Acq On : 9 Dec 2020 1:27
Operator : JC/MD
Sample : L4965-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT014

Quant Time: Dec 09 08:15:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

