

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064998.D
 Acq On : 9 Dec 2020 3:51
 Operator : JC/MD
 Sample : L4970-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4S2-C-TOP

Quant Time: Dec 09 08:25:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	161888	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	311476	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	329525	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	141233	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105296	55.95	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	111.90%	
35) Dibromofluoromethane	7.55	113	97505	52.32	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	104.64%	
50) Toluene-d8	10.06	98	394368	51.79	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	103.58%	
62) 4-Bromofluorobenzene	12.38	95	146763	53.85	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	107.70%	

Target Compounds					Qvalue
16) Acetone	3.78	43	14234	23.007 ug/l	100
20) Methylene Chloride	4.50	84	26939	13.895 ug/l	95
43) Isopropyl Acetate	8.13	43	9491	2.121 ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064998.D
Acq On : 9 Dec 2020 3:51
Operator : JC/MD
Sample : L4970-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4S2-C-TOP

Quant Time: Dec 09 08:25:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

