

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN065000.D
 Acq On : 9 Dec 2020 4:39
 Operator : JC/MD
 Sample : L4970-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4S2-D-TOP

Quant Time: Dec 09 08:28:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	164839	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	317686	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	335321	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136420	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	106946	55.81	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	111.62%	
35) Dibromofluoromethane	7.55	113	97548	51.32	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	102.64%	
50) Toluene-d8	10.06	98	398840	51.35	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	102.70%	
62) 4-Bromofluorobenzene	12.38	95	147286	52.99	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	105.98%	

Target Compounds

					Qvalue	
16) Acetone	3.78	43	11916	18.915	ug/l	99
20) Methylene Chloride	4.51	84	17827	9.031	ug/l	94
43) Isopropyl Acetate	8.13	43	16119	3.532	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN065000.D
Acq On : 9 Dec 2020 4:39
Operator : JC/MD
Sample : L4970-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4S2-D-TOP

Quant Time: Dec 09 08:28:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

