

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN065002.D
 Acq On : 9 Dec 2020 5:27
 Operator : JC/MD
 Sample : L4970-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4S1-A-TOP

Quant Time: Dec 09 08:31:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	161972	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	309283	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	327405	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	137061	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	104575	55.54	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.08%
35) Dibromofluoromethane	7.55	113	97216	52.53	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.06%
50) Toluene-d8	10.06	98	390527	51.65	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.30%
62) 4-Bromofluorobenzene	12.38	95	144827	53.52	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.04%

Target Compounds

					Qvalue
16) Acetone	3.78	43	6140	9.919 ug/l	92
20) Methylene Chloride	4.50	84	7631	3.934 ug/l	92
43) Isopropyl Acetate	8.13	43	16576	3.730 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN065002.D
Acq On : 9 Dec 2020 5:27
Operator : JC/MD
Sample : L4970-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4S1-A-TOP

Quant Time: Dec 09 08:31:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

