

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN065004.D
 Acq On : 9 Dec 2020 6:16
 Operator : JC/MD
 Sample : L4970-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4S1-B-TOP

Quant Time: Dec 09 08:34:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	166237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	324144	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	341676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	142632	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	108304	56.04	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.08%	
35) Dibromofluoromethane		7.55	113	100684	51.91	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.82%	
50) Toluene-d8		10.06	98	407308	51.40	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.80%	
62) 4-Bromofluorobenzene		12.38	95	150670	53.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	6548	10.307	ug/l	95
20) Methylene Chloride	4.50	84	7677	3.856	ug/l	98
43) Isopropyl Acetate	8.13	43	18789	4.035	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN065004.D
Acq On : 9 Dec 2020 6:16
Operator : JC/MD
Sample : L4970-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4S1-B-TOP

Quant Time: Dec 09 08:34:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

