

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN065007.D
 Acq On : 9 Dec 2020 7:28
 Operator : JC/MD
 Sample : L4970-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4S1-C-BOTTOM

Quant Time: Dec 09 08:39:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	163665	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	322005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	340704	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	140771	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.98	65	108286	56.91	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	113.82%	
35) Dibromofluoromethane	7.55	113	99383	51.58	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	103.16%	
50) Toluene-d8	10.06	98	404136	51.33	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	102.66%	
62) 4-Bromofluorobenzene	12.38	95	149520	53.07	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	106.14%	

Target Compounds

					Qvalue
16) Acetone	3.78	43	12134	19.400 ug/l #	87
20) Methylene Chloride	4.51	84	16663	8.502 ug/l	92
43) Isopropyl Acetate	8.13	43	15638	3.380 ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN065007.D
Acq On : 9 Dec 2020 7:28
Operator : JC/MD
Sample : L4970-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4S1-C-BOTTOM

Quant Time: Dec 09 08:39:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

