

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN064968.D
 Acq On : 8 Dec 2020 14:37
 Operator : JC/MD
 Sample : PB133456ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133456ZHE#05

Quant Time: Dec 09 07:15:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	168928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	320994	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	341310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	143749	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	108080	55.03	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	110.06%	
35) Dibromofluoromethane	7.55	113	98926	51.50	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	103.00%	
50) Toluene-d8	10.06	98	406025	51.74	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.37	95	156616	55.76	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	111.52%	

Target Compounds


					Qvalue
16) Acetone	3.78	43	5757	8.917 ug/l	96
20) Methylene Chloride	4.51	84	5404	2.671 ug/l	95
43) Isopropyl Acetate	8.13	43	16575	3.594 ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN064968.D
Acq On : 8 Dec 2020 14:37
Operator : JC/MD
Sample : PB133456ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#05

Quant Time: Dec 09 07:15:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.62 min Scan# 1830

Delta R.T. -0.00 min

Lab File: VN064968.D

Acq: 8 Dec 2020 14:37

Instrument :

MSVOA_N

ClientSampleId :


PB133456ZHE#05

Tot Ion:168 Resp: 168928

Ion Ratio Lower Upper

168 100

99 53.9 43.0 64.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.62

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#16

Acetone

Concen: 8.917 ug/l

RT: 3.78 min Scan# 635

Delta R.T. 0.00 min

Lab File: VN064968.D


Acq: 8 Dec 2020 14:37

Tgt Ion: 43 Resp: 5757

Ion Ratio Lower Upper

43 100

58 29.6 25.7 38.5



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

3.78

2500

2000


1500


1000

500

0

Time--> 3.70 3.75 3.80 3.85









#20
Methvlene Chloride
Concen: 2.671 ug/l
RT: 4.51 min Scan# 860
Delta R.T. 0.01 min
Lab File: VN064968.D
Acq: 8 Dec 2020 14:37

Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#05

Tot Ion: 84 Resp: 5404
Ion Ratio Lower Upper
84 100
49 112.7 93.4 140.0
51 35.6 28.2 42.4
86 56.6 51.7 77.5




Abundance Ion 83.90 (83.60 to 84.60): VN0
Ion 48.90 (48.60 to 49.60): VN0
Ion 50.90 (50.60 to 51.60): VN0
Ion 85.90 (85.60 to 86.60): VN0






#33
1,2-Dichloroethane-d4
Concen: 55.033 ug/l
RT: 7.99 min Scan# 1943
Delta R.T. 0.00 min
Lab File: VN064968.D
Acq: 8 Dec 2020 14:37

Tgt Ion: 65 Resp: 108080
Ion Ratio Lower Upper
65 100
67 53.7 0.0 108.2



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2118

Delta R.T. 0.00 min

Lab File: VN064968.D

Acq: 8 Dec 2020 14:37

Instrument :

MSVOA_N

ClientSampleId :

PB133456ZHE#05


Tot Ion:114 Resp: 320994

Ion Ratio Lower Upper

114 100

63 18.9 0.0 38.6




88 15.7 0.0 32.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 51.505 ug/l

RT: 7.55 min Scan# 1807

Delta R.T. 0.00 min

Lab File: VN064968.D

Acq: 8 Dec 2020 14:37


Tgt Ion:113 Resp: 98926

Ion Ratio Lower Upper

113 100

111 103.2 81.6 122.4



192 20.5 16.2 24.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.594 ug/l

RT: 8.13 min Scan# 1986

Delta R.T. -0.00 min

Lab File: VN064968.D

Acq: 8 Dec 2020 14:37

Instrument :

MSVOA_N

ClientSampleId :

PB133456ZHE#05


Tot Ion: 43 Resp: 16575

Ion Ratio Lower Upper

43 100

61 19.1 17.2 25.8




87 12.2 11.4 17.0



Abundance Ion 43.00 (42.70 to 43.70): VN064968.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN064968.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN064968.D (-1968) (-)



#50

Toluene-d8

Concen: 51.737 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN064968.D


Acq: 8 Dec 2020 14:37

Tgt Ion: 98 Resp: 406025

Ion Ratio Lower Upper



98 100

100 65.1 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN064968.D (-2572) (-)

Ion 100.00 (99.70 to 100.70): VN064968.D (-2572) (-)









#62
4-Bromofluorobenzene
Concen: 55.763 ug/l
RT: 12.37 min Scan# 3307
Delta R.T. -0.00 min
Lab File: VN064968.D
Acq: 8 Dec 2020 14:37




Instrument :
MSVOA_N
ClientSampleId :
PB133456ZHE#05

Tot Ion: 95 Resp: 156616
Ion Ratio Lower Upper
95 100
174 81.3 0.0 160.4
176 78.7 0.0 158.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2997
Delta R.T. -0.00 min
Lab File: VN064968.D
Acq: 8 Dec 2020 14:37

Tgt Ion: 117 Resp: 341310
Ion Ratio Lower Upper
117 100
82 55.0 44.2 66.2
119 33.1 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3600

Delta R.T. -0.00 min

Lab File: VN064968.D

Acq: 8 Dec 2020 14:37

Instrument :

MSVOA_N

ClientSampleId :

PB133456ZHE#05


Tot Ion:152 Resp: 143749

Ion Ratio Lower Upper

152 100

115 59.3 29.7 89.1

150 157.2 0.0 350.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

