

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120820\
 Data File : VN065007.D
 Acq On : 9 Dec 2020 7:28
 Operator : JC/MD
 Sample : L4970-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 4S1-C-BOTTOM

Quant Time: Dec 09 08:39:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	163665	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	322005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	340704	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	140771	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.98	65	108286	56.91	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.82%	
35) Dibromofluoromethane		7.55	113	99383	51.58	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.16%	
50) Toluene-d8		10.06	98	404136	51.33	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.66%	
62) 4-Bromofluorobenzene		12.38	95	149520	53.07	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.14%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	12134	19.400	ug/l #	87
20) Methylene Chloride	4.51	84	16663	8.502	ug/l	92
43) Isopropyl Acetate	8.13	43	15638	3.380	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120820\
Data File : VN065007.D
Acq On : 9 Dec 2020 7:28
Operator : JC/MD
Sample : L4970-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
4S1-C-BOTTOM

Quant Time: Dec 09 08:39:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN065007.D

Acq: 9 Dec 2020 7:28


Instrument :
MSVOA_N
ClientSampleId :
4S1-C-BOTTOM

Tot Ion:168 Resp: 163665

Ion Ratio Lower Upper

168 100

99 53.9 43.0 64.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.63

60000



40000

20000

0

7.55 7.60 7.65 7.70

Time-->



#16

Acetone

Concen: 19.400 ug/l

RT: 3.78 min Scan# 633

Delta R.T. -0.00 min

Lab File: VN065007.D


Acq: 9 Dec 2020 7:28

Tgt Ion: 43 Resp: 12134

Ion Ratio Lower Upper

43 100

58 25.1 25.7 38.5#



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN

3.78

5000

4000

3000


2000


1000

0

3.70 3.75 3.80 3.85 3.90

Time-->









#20
Methylene Chloride
Concen: 8.502 ug/l
RT: 4.51 min Scan# 860
Delta R.T. 0.01 min
Lab File: VN065007.D
Acq: 9 Dec 2020 7:28




Instrument :
MSVOA_N
ClientSampleId :
4S1-C-BOTTOM


Tot Ion: 84 Resp: 16663
Ion Ratio Lower Upper
84 100
49 107.3 93.4 140.0
51 34.1 28.2 42.4
86 56.3 51.7 77.5



#33
1,2-Dichloroethane-d4
Concen: 56.911 ug/l
RT: 7.98 min Scan# 1942
Delta R.T. -0.00 min
Lab File: VN065007.D
Acq: 9 Dec 2020 7:28

Tgt Ion: 65 Resp: 108286
Ion Ratio Lower Upper
65 100
67 53.6 0.0 108.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.55 min Scan# 2118
Delta R.T. 0.00 min
Lab File: VN065007.D
Acq: 9 Dec 2020 7:28

Instrument :
MSVOA_N
ClientSampleId :
4S1-C-BOTTOM

Tot Ion	Ratio	Lower	Upper
114	100		
63	18.4	0.0	38.6
88	15.2	0.0	32.6




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): V
Ion 87.90 (87.60 to 88.60): V






#35
Dibromofluoromethane
Concen: 51.580 ug/l
RT: 7.55 min Scan# 1807
Delta R.T. 0.00 min
Lab File: VN065007.D
Acq: 9 Dec 2020 7:28

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.5	81.6	122.4
192	19.9	16.2	24.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.380 ug/l

RT: 8.13 min Scan# 1987

Delta R.T. 0.00 min

Lab File: VN065007.D

Acq: 9 Dec 2020 7:28

Instrument :

MSVOA_N

ClientSampleId :

4S1-C-BOTTOM


Tot Ion: 43 Resp: 15638

Ion Ratio Lower Upper

43 100

61 16.0 17.2 25.8#




87 10.8 11.4 17.0#



Abundance Ion 43.00 (42.70 to 43.70): VN064779.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN064779.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN064779.D (-1968) (-)



#50

Toluene-d8

Concen: 51.334 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN065007.D


Acq: 9 Dec 2020 7:28

Tgt Ion: 98 Resp: 404136

Ion Ratio Lower Upper



98 100

100 64.4 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN064779.D (-2572) (-)

Ion 100.00 (99.70 to 100.70): VN064779.D (-2572) (-)









#62
4-Bromofluorobenzene
Concen: 53.069 ug/l
RT: 12.38 min Scan# 3308
Delta R.T. 0.00 min
Lab File: VN065007.D
Acq: 9 Dec 2020 7:28




Instrument :
MSVOA_N
ClientSampleId :
4S1-C-BOTTOM

Tot Ion: 95 Resp: 149520
Ion Ratio Lower Upper
95 100
174 82.6 0.0 160.4
176 79.9 0.0 158.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN065007.D
Acq: 9 Dec 2020 7:28

Tgt Ion: 117 Resp: 340704
Ion Ratio Lower Upper
117 100
82 53.2 44.2 66.2
119 31.8 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3601

Delta R.T. 0.00 min

Lab File: VN065007.D

Acq: 9 Dec 2020 7:28

Instrument :

MSVOA_N

ClientSampleId :

4S1-C-BOTTOM


Tot Ion:152 Resp: 140771

Ion Ratio Lower Upper

152 100

115 58.3 29.7 89.1

150 157.6 0.0 350.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

