

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069857.D
 Acq On : 8 Dec 2021 13:31
 Operator : JC/MD
 Sample : M4957-01DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW17811-20211206DL

Quant Time: Dec 09 11:38:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1204860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2188539	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2310447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	856048	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	941703	50.62	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.24%			
35) Dibromofluoromethane	8.024	113	650688	47.47	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 94.94%			
50) Toluene-d8	10.440	98	2832730	51.46	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.92%			
62) 4-Bromofluorobenzene	12.731	95	1069769	53.65	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 107.30%			
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	4.189	96	6847	0.60	ug/l	95
27) cis-1,2-Dichloroethene	7.327	96	127654	8.48	ug/l	96
44) Trichloroethene	9.217	130	1302523	85.69	ug/l	89
45) 1,2-Dichloropropane	9.491	63	80613	4.71	ug/l	99
64) Tetrachloroethene	10.977	164	58737	3.75	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069857.D
Acq On : 8 Dec 2021 13:31
Operator : JC/MD
Sample : M4957-01DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW17811-20211206DL

Quant Time: Dec 09 11:38:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

