

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069859.D
 Acq On : 8 Dec 2021 14:22
 Operator : JC/MD
 Sample : M4957-04DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW179D2-20211206DL

Quant Time: Dec 09 11:39:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1272819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2304370	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2465577	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	920470	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1005035	51.141	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.280%			
35) Dibromofluoromethane	8.021	113	690782	47.858	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 95.720%			
50) Toluene-d8	10.440	98	2989766	51.585	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.180%			
62) 4-Bromofluorobenzene	12.731	95	1149586	54.751	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 109.500%			
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.329	96	91765	5.768	ug/l	95
44) Trichloroethene	9.217	130	709131	44.309	ug/l	91
64) Tetrachloroethene	10.977	164	49472	2.959	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069859.D
Acq On : 8 Dec 2021 14:22
Operator : JC/MD
Sample : M4957-04DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW179D2-20211206DL

Quant Time: Dec 09 11:39:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

