

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069860.D
 Acq On : 8 Dec 2021 14:47
 Operator : JC/MD
 Sample : M4956-03DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D3-20211202DL

Quant Time: Dec 09 11:39:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1360259	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2471633	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2671890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1005392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1061250	50.530	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.060%
35) Dibromofluoromethane	8.024	113	731679	47.261	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.520%
50) Toluene-d8	10.440	98	3219405	51.789	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.580%
62) 4-Bromofluorobenzene	12.731	95	1250894	55.544	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.080%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.213	101	572541	42.581	ug/l	96
44) Trichloroethene	9.220	130	232851	13.565	ug/l	93
64) Tetrachloroethene	10.977	164	125056	6.903	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069860.D
Acq On : 8 Dec 2021 14:47
Operator : JC/MD
Sample : M4956-03DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D3-20211202DL

Quant Time: Dec 09 11:39:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

