

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069862.D
 Acq On : 8 Dec 2021 15:38
 Operator : JC/MD
 Sample : M4956-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-20211203

Quant Time: Dec 09 11:40:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1292656	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2349672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2559861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	992683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1023610	51.287	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 102.580%		
35) Dibromofluoromethane	8.024	113	709269	48.191	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 96.380%		
50) Toluene-d8	10.440	98	3067123	51.900	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 103.800%		
62) 4-Bromofluorobenzene	12.731	95	1204963	56.282	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 112.560%		
Target Compounds						
16) Acetone	4.296	43	70094	5.838	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069862.D
Acq On : 8 Dec 2021 15:38
Operator : JC/MD
Sample : M4956-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-20211203

Quant Time: Dec 09 11:40:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

