

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069863.D
 Acq On : 8 Dec 2021 16:04
 Operator : JC/MD
 Sample : M4956-22
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB02-20211206

Quant Time: Dec 09 11:41:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1301162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2346737	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2548832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	968881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1025055	51.023	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.040%
35) Dibromofluoromethane	8.021	113	704285	47.912	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.820%
50) Toluene-d8	10.443	98	3059817	51.841	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.680%
62) 4-Bromofluorobenzene	12.733	95	1178539	55.116	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.240%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	60512	5.007	ug/l	97
20) Methylene Chloride	5.098	84	17465	1.087	ug/l	88
68) m/p-Xylenes	11.950	106	16974	0.496	ug/l	98
69) o-Xylene	12.280	106	11155	0.324	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069863.D
Acq On : 8 Dec 2021 16:04
Operator : JC/MD
Sample : M4956-22
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB02-20211206

Quant Time: Dec 09 11:41:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

