

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069864.D
 Acq On : 8 Dec 2021 16:29
 Operator : JC/MD
 Sample : M4956-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP07-20211205

Quant Time: Dec 09 11:41:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1344855	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2433880	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2631401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	985293	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1050874	50.61	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.22%
35) Dibromofluoromethane	8.021	113	733707	48.13	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.26%
50) Toluene-d8	10.440	98	3158466	51.60	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.20%
62) 4-Bromofluorobenzene	12.733	95	1221217	55.07	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.14%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.075	85	105402	6.72	ug/l	98
7) Trichlorofluoromethane	3.381	101	2428457	102.40	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.213	101	25051	1.88	ug/l #	63
12) 1,1-Dichloroethene	4.186	96	394723	30.81	ug/l	90
24) 1,1-Dichloroethane	6.391	63	78928	2.70	ug/l	99
27) cis-1,2-Dichloroethene	7.324	96	29458	1.75	ug/l	94
30) Chloroform	7.817	83	90455	3.06	ug/l	99
32) 1,1,1-Trichloroethane	8.016	97	89759	3.57	ug/l #	93
44) Trichloroethene	9.217	130	957315	56.63	ug/l	90
64) Tetrachloroethene	10.977	164	799248	44.79	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069864.D
Acq On : 8 Dec 2021 16:29
Operator : JC/MD
Sample : M4956-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP07-20211205

Quant Time: Dec 09 11:41:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

