

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069866.D
 Acq On : 8 Dec 2021 17:20
 Operator : JC/MD
 Sample : M4957-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW179D1-20211206

Quant Time: Dec 09 11:42:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1335328	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2409197	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2627180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	986289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1030391	49.98	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.96%
35) Dibromofluoromethane	8.024	113	727881	48.23	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.46%
50) Toluene-d8	10.440	98	3160380	52.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.32%
62) 4-Bromofluorobenzene	12.731	95	1224306	55.77	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.54%
Target Compounds						
27) cis-1,2-Dichloroethene	7.329	96	142782	8.55	ug/l	95
44) Trichloroethene	9.217	130	1190159	71.13	ug/l	90
64) Tetrachloroethene	10.979	164	103634	5.82	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069866.D
Acq On : 8 Dec 2021 17:20
Operator : JC/MD
Sample : M4957-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW179D1-20211206

Quant Time: Dec 09 11:42:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

