

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069870.D
 Acq On : 8 Dec 2021 19:02
 Operator : JC/MD
 Sample : PB141231ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141231ZHE#01

Quant Time: Dec 09 11:45:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1257837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2316230	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2555939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	990238	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1007934	51.899	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.800%			
35) Dibromofluoromethane	8.024	113	697111	48.049	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.100%			
50) Toluene-d8	10.441	98	3057629	52.486	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.980%			
62) 4-Bromofluorobenzene	12.731	95	1212226	57.438	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 114.880%			
Target Compounds						
16) Acetone	4.293	43	94250	8.068	ug/l	99
18) Methyl Acetate	4.870	43	12797	0.363	ug/l #	68
20) Methylene Chloride	5.098	84	149073	9.599	ug/l #	84
25) 2-Butanone	7.346	43	17715	1.116	ug/l #	89
44) Trichloroethene	9.215	130	8435	0.524	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069870.D
Acq On : 8 Dec 2021 19:02
Operator : JC/MD
Sample : PB141231ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141231ZHE#01

Quant Time: Dec 09 11:45:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

