

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069871.D
 Acq On : 8 Dec 2021 19:27
 Operator : JC/MD
 Sample : PB141231ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141231ZHE#02

Quant Time: Dec 09 11:45:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1308588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2404541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2627635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1043061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1033648	51.159	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.320%			
35) Dibromofluoromethane	8.024	113	710814	47.194	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 94.380%			
50) Toluene-d8	10.440	98	3146228	52.024	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.040%			
62) 4-Bromofluorobenzene	12.731	95	1235099	56.373	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 112.740%			
Target Compounds						
16) Acetone	4.291	43	152860	12.577	ug/l	95
18) Methyl Acetate	4.862	43	25818	0.704	ug/l #	71
20) Methylene Chloride	5.095	84	37482	2.320	ug/l	86
25) 2-Butanone	7.348	43	25840	1.565	ug/l	92
43) Isopropyl Acetate	8.563	43	163427	3.241	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069871.D
Acq On : 8 Dec 2021 19:27
Operator : JC/MD
Sample : PB141231ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141231ZHE#02

Quant Time: Dec 09 11:45:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

