

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069882.D
 Acq On : 9 Dec 2021 00:57
 Operator : JC/MD
 Sample : PB141231ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141231ZHE#07

Quant Time: Dec 09 11:53:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1198186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2209323	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2419429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	915253	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	945739	51.121	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.240%	
35) Dibromofluoromethane	8.024	113	651369	47.069	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 94.140%	
50) Toluene-d8	10.440	98	2903035	52.244	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.480%	
62) 4-Bromofluorobenzene	12.731	95	1137210	56.491	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 112.980%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	102396	9.201	ug/l	99
18) Methyl Acetate	4.859	43	22137	0.659	ug/l #	59
20) Methylene Chloride	5.101	84	47371	3.202	ug/l	85
25) 2-Butanone	7.343	43	22134	1.464	ug/l #	89
43) Isopropyl Acetate	8.563	43	170504	3.681	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069882.D
Acq On : 9 Dec 2021 00:57
Operator : JC/MD
Sample : PB141231ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141231ZHE#07

Quant Time: Dec 09 11:53:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

