

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069885.D
 Acq On : 9 Dec 2021 2:13
 Operator : JC/MD
 Sample : PB141231ZHE#10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141231ZHE#10

Quant Time: Dec 09 11:54:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1378840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2540855	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2815123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1110560	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1064676	50.010	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.020%
35) Dibromofluoromethane	8.024	113	746398	46.898	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.800%
50) Toluene-d8	10.440	98	3329901	52.107	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.220%
62) 4-Bromofluorobenzene	12.731	95	1336405	57.724	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.440%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	86573	6.760	ug/l	98
18) Methyl Acetate	4.862	43	44562	1.153	ug/l	95
20) Methylene Chloride	5.101	84	44068	2.588	ug/l	91
25) 2-Butanone	7.348	43	19399	1.115	ug/l	92
43) Isopropyl Acetate	8.563	43	213144	4.001	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069885.D
Acq On : 9 Dec 2021 2:13
Operator : JC/MD
Sample : PB141231ZHE#10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141231ZHE#10

Quant Time: Dec 09 11:54:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

