

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069889.D
 Acq On : 9 Dec 2021 3:55
 Operator : JC/MD
 Sample : PB141231TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141231TB

Quant Time: Dec 09 11:57:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1323558	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2474043	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2720506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1099500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1050434	51.402	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.800%	
35) Dibromofluoromethane	8.024	113	729837	47.096	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 94.200%	
50) Toluene-d8	10.440	98	3228573	51.885	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.780%	
62) 4-Bromofluorobenzene	12.731	95	1310373	58.128	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 116.260%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	150220	12.220	ug/l	99
18) Methyl Acetate	4.862	43	60469	1.630	ug/l #	73
20) Methylene Chloride	5.095	84	36550	2.237	ug/l #	83
25) 2-Butanone	7.340	43	24875	1.489	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069889.D
Acq On : 9 Dec 2021 3:55
Operator : JC/MD
Sample : PB141231TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141231TB

Quant Time: Dec 09 11:57:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

