

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069890.D
 Acq On : 9 Dec 2021 4:21
 Operator : JC/MD
 Sample : M4958-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 D-01

Quant Time: Dec 09 11:57:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1281668	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2398967	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2642133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1090720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1021061	51.597	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.200%			
35) Dibromofluoromethane	8.024	113	718105	47.789	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 95.580%			
50) Toluene-d8	10.440	98	3117805	51.673	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.340%			
62) 4-Bromofluorobenzene	12.731	95	1276724	58.408	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 116.820%			
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	105870	8.894	ug/l	99
18) Methyl Acetate	4.862	43	63936	1.779	ug/l	94
20) Methylene Chloride	5.095	84	31693	2.003	ug/l	88
25) 2-Butanone	7.348	43	20341	1.258	ug/l	96
43) Isopropyl Acetate	8.563	43	179181	3.562	ug/l #	81
52) Toluene	10.507	92	45474	1.067	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069890.D
Acq On : 9 Dec 2021 4:21
Operator : JC/MD
Sample : M4958-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
D-01

Quant Time: Dec 09 11:57:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

