

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069894.D
 Acq On : 9 Dec 2021 6:02
 Operator : JC/MD
 Sample : M4959-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3S

Quant Time: Dec 09 11:59:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1261446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2404728	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2667981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1070191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1012080	51.964	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.920%	
35) Dibromofluoromethane	8.024	113	707091	46.943	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 93.880%	
50) Toluene-d8	10.440	98	3120070	51.587	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.180%	
62) 4-Bromofluorobenzene	12.731	95	1283698	58.586	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 117.180%	
Target Compounds						
						Qvalue
16) Acetone	4.296	43	97289	8.304	ug/l	90
18) Methyl Acetate	4.870	43	54117	1.530	ug/l	90
20) Methylene Chloride	5.103	84	35366	2.271	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069894.D
Acq On : 9 Dec 2021 6:02
Operator : JC/MD
Sample : M4959-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3S

Quant Time: Dec 09 11:59:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

