

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
 Data File : VN069900.D
 Acq On : 9 Dec 2021 8:35
 Operator : JC/MD
 Sample : M4965-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BLUE-CONCRETE-COMP

Quant Time: Dec 09 12:03:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1246425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2338213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2644833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1093657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	1008931	52.426	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.860%
35) Dibromofluoromethane	8.024	113	724693	49.480	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.960%
50) Toluene-d8	10.440	98	3105956	52.815	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.620%
62) 4-Bromofluorobenzene	12.731	95	1282300	60.187	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	120.380%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	106163	9.170	ug/l	99
18) Methyl Acetate	4.867	43	43961	1.258	ug/l #	73
20) Methylene Chloride	5.100	84	42442	2.758	ug/l	87
43) Isopropyl Acetate	8.563	43	188956	3.854	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120821\
Data File : VN069900.D
Acq On : 9 Dec 2021 8:35
Operator : JC/MD
Sample : M4965-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BLUE-CONCRETE-COMP

Quant Time: Dec 09 12:03:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

