

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075638.D
 Acq On : 08 Dec 2022 10:20
 Operator : JC\MD
 Sample : VN1208MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1208MBL01

Quant Time: Dec 09 02:03:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	221502	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	363769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	323693	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	143969	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	126140	106.090	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	212.180%#
35) Dibromofluoromethane	8.177	113	113821	79.651	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	159.300%#
50) Toluene-d8	10.571	98	418562	130.048	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	260.100%#
62) 4-Bromofluorobenzene	12.853	95	135938	57.222	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075638.D
Acq On : 08 Dec 2022 10:20
Operator : JC\MD
Sample : VN1208MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1208MBL01

Quant Time: Dec 09 02:03:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

