

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075645.D
 Acq On : 08 Dec 2022 14:08
 Operator : JC\MD
 Sample : N5944-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW-BR-04-95-105-120722

Quant Time: Dec 09 02:07:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	166553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	275840	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	237917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103155	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	45264	50.629	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.260%
35) Dibromofluoromethane	8.171	113	60145	55.505	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.020%
50) Toluene-d8	10.571	98	119772	49.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.160%
62) 4-Bromofluorobenzene	12.853	95	86322	47.920	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.840%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.353	96	3426	2.221	ug/l #	61
16) Acetone	4.453	43	2682	3.216	ug/l #	74
27) cis-1,2-Dichloroethene	7.494	96	21281	10.707	ug/l	83
44) Trichloroethene	9.359	130	54818	27.690	ug/l	93
52) Toluene	10.635	92	77736	16.673	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075645.D
Acq On : 08 Dec 2022 14:08
Operator : JC\MD
Sample : N5944-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW-BR-04-95-105-120722

Quant Time: Dec 09 02:07:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

