

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075647.D
 Acq On : 08 Dec 2022 14:56
 Operator : JC\MD
 Sample : N5924-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF1222

Quant Time: Dec 09 02:07:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	165635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	269600	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	232593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97140	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	44641	50.209	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.420%
35) Dibromofluoromethane	8.177	113	57385	54.184	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.360%
50) Toluene-d8	10.571	98	113615	47.631	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.260%
62) 4-Bromofluorobenzene	12.853	95	79216	44.993	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075647.D
Acq On : 08 Dec 2022 14:56
Operator : JC\MD
Sample : N5924-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF1222

Quant Time: Dec 09 02:07:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

