

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
 Data File : VN075648.D
 Acq On : 08 Dec 2022 15:20
 Operator : JC\MD
 Sample : N5924-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MID1222

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/09/2022
 Supervised By : Mahesh Dadoda 12/09/2022

Quant Time: Dec 09 02:08:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 23 00:18:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	164129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	268597	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	234065	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100998	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	39057	44.332	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.660%		
35) Dibromofluoromethane	8.177	113	50592	47.948	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.900%		
50) Toluene-d8	10.571	98	99565	41.896	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	83.800%#		
62) 4-Bromofluorobenzene	12.853	95	72400	41.275	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	82.560%#		
Target Compounds						Qvalue
11) Tert butyl alcohol	5.530	59	6632m	17.932	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120822\
Data File : VN075648.D
Acq On : 08 Dec 2022 15:20
Operator : JC\MD
Sample : N5924-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MID1222

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/09/2022
Supervised By :Mahesh Dadoda 12/09/2022

Quant Time: Dec 09 02:08:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112122W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 23 00:18:03 2022
Response via : Initial Calibration

